

LASER SYSTEM FOR THICKNESS CONTROL AND CUTTING OF BIOMATERIALS

User's manual

22, Logoisky tract, Minsk
220090, Republic of Belarus
tel/fax: +375 17 357 36 57
info@riftek.com
www.riftek.com

Contents

1.	Safety precautions.....	3
2.	CE compliance.....	3
3.	General information.....	3
4.	Structure and operating principle.....	3
4.1.	Structure	3
4.2.	Operating principle.....	4
5.	Basic technical data.....	4
6.	Software	4
6.1.	Main window.....	4
6.2.	Program settings.....	5
6.3.	Color palette.....	6
7.	Intended use.....	7
7.1.	Preparation for work.....	7
7.2.	Calibration.....	7
7.3.	Scanning.....	8
7.4.	Importing and placing patterns.....	8
7.5.	Cutting	9
8.	Warranty policy.....	9
9.	Revisions	9
10.	Distributors.....	10

1. Safety precautions

- Only persons who have studied this User's Manual are allowed to work with the system.
- Use the supply voltage and interfaces indicated in the specification.
- When connecting/disconnecting cables, the system must be powered off.

2. CE compliance

The system has been developed for use in industry and meets the requirements of the following Directives:

- EU directive 2014/30/EU. Electromagnetic compatibility (EMC).
- EU directive 2011/65/EU, "RoHS" category 9.

3

3. General information

The system is designed for non-contact automated thickness control and cutting of biomaterials used in cardiac and vascular surgery.

The main functions of the system:

- 1) Measuring the distribution of the thickness of the biomaterial over the area and building a height map.
- 2) Optimal placement of product patterns on the height map of the biomaterial in accordance with the specified thickness.
- 3) Laser cutting of biomaterial in accordance with patterns.

4. Structure and operating principle

4.1. Structure

The structural scheme of the system is shown in Figure 1. The system is implemented on the basis of a CNC laser cutting machine with a Ruida controller or similar. The machine is controlled via a central computer using specialized software. The 2D laser scanner is mounted on the movement system of the machine. In the measurement mode, the CNC system of the machine moves the sensor line by line (snake) over the material. The sensor measures the distance (Z coordinate) to the surface of the material. Data pickup from the sensor is synchronized with its movement (XY coordinates) and the result is transmitted to the computer via the Ethernet port. The software creates a material height map (X,Y,Z).




Figure 1 - Structural scheme

4.2. Operating principle

The operation of the system includes the following stages:

1. Calibration. At this stage, the system scans the surface of the object plate and generates a height map in the scanner coordinate system.
2. Measurement of the biomaterial thickness and formation of its height map. At this stage, the system scans the surface of the biomaterial placed on the object plate. The thickness of the biomaterial is calculated as the difference between the heights of surface of the biomaterial and the object plate.
3. Programmatic placement of patterns selected from the database on the height map of the biomaterial.
4. Laser cutting of the biomaterial in accordance with the location of the patterns.

4

5. Basic technical data


Parameter	Value
Measurement object	pericardial tissue for making prostheses
Measured parameter	thickness distribution
Thickness measurement error, mm	±0.03
Scan area size	determined by the model of the laser cutting machine
Scan speed, up to, mm/s	100
Power supply	Three-phase AC network with frequency (50±1) Hz, rated voltage 220/380 V with voltage tolerance ±10%
Template file format	.dxf
2D laser scanner	RF627.Blue-25/10-8/11
Operating frequency, profiles/s	1000

6. Software

6.1. Main window

To start working with the system, go to the software folder and run **LaserNC.exe**.

The main window:




Description of interface elements:

#	Element	Description
	Interface language	Language selection button.
	Patterns import	Pattern file import button.
	Carriage movement	Buttons for controlling the manual movement of the carriage.
	Scanning speed	This button opens a panel for entering the movement speed.
	Start cutting	Start cutting the biomaterial in accordance with the pattern.
	Stop cutting	Stop cutting the biomaterial.
1	-	Area for building a height map and placing patterns.
2	-	Thickness color scale.
3	-	Software settings.
	Remove all patterns from the workspace	Remove all imported patterns from the working area, keeping the height map.
	Reset cutting map	Remove the cutting map from the working area.
	Home	Return the carriage to its starting position.
	Scanner calibration	Start the calibration process.
	Start scanning	Start the scanning process in the specified area.

5

6.2. Program settings

To open the program settings, click . The **Program Settings** window will appear. To apply the specified settings, it is necessary to click the **Apply** button. If no changes are required, click the **Cancel** button.



Description of parameters:

Name	Description
Cutting X, Cutting Y	These values are set to correct the starting position of laser cutting along the corresponding axes. They can be used when it is necessary to adjust the cutting coordinates due to possible deformations of the material or equipment errors. Correction coefficients are specified in millimeters and are usually set depending on the characteristics of the material and machine settings.
Thickness	This value is specified in millimeters and is used to correct the scan data. This coefficient can be added to the resulting data to take into account the possible measurement error caused by various factors, such as material deformation and penetration of laser radiation into the material.
Machine address	The network address of the machine controller.
Machine port	The network port of the machine controller.
Scanner S/N	The unique address of the laser scanner. It is used to search for and connect to the scanner, as well as for software licensing.
Step size (mm)*	The number of millimeters per pulse of the machine stepper motor driver. This parameter is used for correct rendering of the final result*.
Laser power (%)	The laser power of the cutting machine. Values: from 1 to 100%.
Laser speed (mm/s)	The speed at which the carriage moves when cutting the material.
Calibration file	Selecting the calibration file.
Pattern setting	Selecting the color of the imported pattern.


* In most cases, this parameter is unique for each machine. It is set only once when changing the machine or machine equipment.

6.3. Color palette

The color palette makes it possible to split the thickness of the material into color layers when building its height map.

To adjust the color palette, double-click on the thickness scale in the working area of the main window. After setting the parameters, click the **Apply** button. If no changes are required, click the **Cancel** button.

7



7. Intended use

7.1. Preparation for work

1. Check if the machine is turned on.
2. Start the software.
3. Wait for the program to find and connect to devices.
4. After successful connection, you can start working with the system.

NOTE. If the scanner or machine is not available for connection, a message will be displayed stating that one of the devices is unavailable.

7.2. Calibration

Before starting the calibration, make sure that the carriage is in the starting position and the object plate is installed in the machine. Start the calibration by clicking the **SCANNER CALIBRATION** button. During calibration, the scanner passes over the entire calibration plate. After the calibration is completed, the program will display a message about its success or failure. If successful, the calibration files will be saved in the program folder with a name containing the date and time of their creation.

An example of calibration files:


-  dump_19_12_2022_10_24.bin
-  dump_19_12_2022_10_24_profiles

7.3. Scanning

To scan the biomaterial, follow these steps:

- Make sure that the machine carriage is in the starting position.
- Place the biomaterial on the object plate and straighten it carefully.
- Select the scan area with the mouse. For convenience, the grid on the screen corresponds to the grid on the surface of the object plate.
- Click the **START SCANNING** button to start the scanning process.
- Scanning will be performed within the selected area.
- If necessary, the scanning process can be repeated.

After the successful completion of the scanning process, a biomaterial height map will be built in the working area of the program. To control the thickness at a specific point, you need to move the cursor over it. The thickness value will be displayed next to the cursor.



7.4. Importing and placing patterns

To import a pattern, do the following:


- Click the **PATTERNS IMPORT** button.
- In the window that appears, select the required pattern in **.dxf** format. The selected pattern will be placed in the working area.

For optimal placement of the pattern on the height map, the following operations can be performed:

- To move, select a part with one click, and then hold down the left mouse button or use the arrows on the keyboard.
- To delete, press the **Delete** key on the keyboard.
- To copy and paste, call the context menu by pressing the right mouse button on the pattern, or use the hotkeys (**Ctrl+C** – copy, **Ctrl+V** – paste).

- To rotate, call the context menu by pressing the right mouse button, and then specify the rotation angle in the input field, or rotate the pattern with the mouse wheel.

NOTE. Patterns must be formed as broken lines, since the laser cutting machine only accepts data in this form. The number of patterns for import is not limited.



7.5. Cutting

After placing the patterns on the height map of the biomaterial, the cutting operation can be performed. Before cutting, make sure of the following:

- The water cooling system of the cutting laser is turned on and working properly.
- Blowing is turned on.
- The machine cover is closed.

To start cutting, click the **START CUTTING** button.

The system will cut the material in accordance with the patterns.

To forcibly stop cutting, click the **STOP CUTTING** button.

8. Warranty policy

Warranty assurance – 24 months from the date of putting in operation; warranty shelf-life – 12 months.

9. Revisions

Date	Revision	Description
10.04.2023	1.0.0	Starting document.

10. Distributors

AUSTRALIA

Applied Measurement Australia Pty Ltd
RAILWAY INSTRUMENTS ONLY
Thornton Plaza, Unit 5,
27 Thornton Crescent, Mitcham
VIC 3132, Australia
Tel: +61 39874 5777
Fax: +61 39874 5888
sales@appliedmeasurement.com.au
www.appliedmeasurement.com.au

BRAZIL

CAPI Controle e Automacao Ltda
Rua Itororo, 121, CEP 13466-240
Americana-SP, Brazil
Tel: +55 19 36047068
Fax: +55 19 34681791
capi@capicontrole.com.br
www.capicontrole.com.br

CHINA

Beijing Gemston Mechanical & Electrical Equipment Co., Ltd
Room 613, Anfu Mansion, Fengtai District, Beijing, China
Tel: +86 10 6765 0516
Fax: +86 10 6765 6966
Mobile: +86 137 1755 1423
dh0526@163.com
www.baof.cn

CHINA

Zhenshangyou Technologies Co., Ltd.
Rm 2205-2210, Zhongyou Hotel
1110 Nanshan Road, Nanshan District 518054 Shenzhen, China
Tel: +86 755-26528100/8011/8012
Fax: +86 755-26528210/26435640
info@51sensors.com
www.51sensors.com

ESTONIA

RD Resolve OU
Punane 14a-304
Tallinn, 13619, Estonia
Tel: +3726206506
rdresolv@gmail.com

BENELUX

Althen Sensors & Controls BV
Vlietweg 17a, 2266KA Leidschendam
The Netherlands
Tel: +31 70 3924421
Fax: +31 70 3644249
sales@althen.nl
www.althensensors.com
www.althencontrols.com

BULGARIA

MTL ASCO Sp. z o.o.
RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120 Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

CHINA

Xi'an Win-Success Automation Technology Co.,Ltd
Room 3-1-1039, Iduhui Building, No.11 Tangyan South Road High-Tech Zone, Xi'an Shaanxi PRC, China
Tel: +86-29-81106280
Fax: +86-29-81106285
Mob: +86-133-1927-1405
info@maxsenor.com
www.maxsensor.com

CROATIA,

CZECH REPUBLIC

MTL ASCO Sp. z o.o.
RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120 Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

FINLAND

Kvalitest Industrial Oy
EXCEPT FOR RAILWAY INSTRUMENTS
Punasillantie 31 A, 40950 Muurame, Finland
Tel: +358 0 20 730 6070
sales@kvalitest.com
www.kvalitest.com
www.kvalitest.fi

BOSNIA AND HERZEGOVINA

MTL ASCO Sp. z o.o.
RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120 Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

CHILE

Verne SpA
Apoquindo 2818, oficina 31 Las Condes, Santiago, Chile
Tel: +56 2 228858633
info@verne.cl
jsaavedra@verne.cl
www.verne.cl

CHINA

Micron-Metrology co., Ltd
No.2, Kecheng Rd., Industrial Park District, Suzhou, Jiangsu Province, China
Tel: +86 189 1806 9807
sales@micron-metrology.cn
www.micron-metrology.cn

DENMARK

BLConsult
Ryssbalt 294
95 291 Kalix, Sweden
Tel: +46 70 663 19 25
info@blconsult.se
www.blconsult.se

FINLAND

TERASPYORA-STEELWHEEL OY
RAILWAY INSTRUMENTS ONLY
Juvan teollisuuskatu 28 FI-02920 ESPOO, Finland
Tel: +358 400 422 900
Fax: +358 9 2511 5510
steelwheel@steelwheel.fi
www.teraspyora.fi

FRANCE**BLET Measurement Group S.A.S.**

1 avenue du President Georges
Pompidou, 92500 Rueil
Malmaison, France
Tel: + 33(0)1 80 88 57 85
Fax: +33(0)1 80 88 57 93
technique@blet-mesure.fr
www.blet-mesure.fr

GERMANY**Disynet GmbH**

Breyeller Str. 2, 41379
Brueggen, Germany
Tel: +49 2157 8799-0
Fax: +49 2157 8799-22
disynet@sensoren.de
www.sensoren.de

GERMANY**Finger GmbH & Co. KG
OPTICAL MICROMETERS ONLY**

Sapelloh 172, 31606
Warmsen, Germany
Tel: +49 5767 96020
Fax: +49 5767 93004
finger@finger-kg.de
www.finger-kg.de

GERMANY**Hylewicz CNC-Technik
SHTRIKH-2 ONLY**

Siemensstrasse 13-15
47608 Geldern, Germany
Tel: +49 2831 91021-20
Fax: +49 2831 91021-99
info@cnc-step.de
www.cnc-step.de

GERMANY**ALTHEN GmbH MeЯ- und
Sensortechnik**

Dieselstrasse 2, 65779
Kelkheim, Germany
Tel: +49 0 6195 7 00 60
info@althen.de
www.althensensors.com/de/

HUNGARY**MTL ASCO Sp. z o.o.**

RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120
Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

INDIA**Influx Big Data Solutions Pvt
Ltd**

No:2, Krishvi, Ground Floor,
Old Airport Road, Domlur,
Bangalore - 560071, India
Tel: +91 73 37748490
Tel: +91 94 48492380
milan@influxtechnology.com
support_india@influxtechnology.com
www.influxtechnology.com

INDIA**Paragon Instrumentation
Engineers Pvt. Ltd.**

RAILWAY INSTRUMENTS ONLY
200, Station Road,
Roorkee, 247 667, India
Tel: +91-1332-272394
tanuj@paragoninstruments.com
www.paragoninstruments.com

INDONESIA**PT. DHAYA BASWARA
SANIYASA**

Botanic Junction Blok H-9 NO. 7
Mega Kebon Jeruk, Joglo
Jakarta 11640, Indonesia
Tel: + 62 21 29325859
management@ptdbs.co.id

IRAN**Novin Industrial
Development Grp.**

Tel: +98 21 44022093-6
Fax: +98 21 43858794
Mobile: +98 9123207518
info@novnid.com
www.novnid.com

ISRAEL**Nisso Dekalo Import
Export LTD**

1 David Hamelech Street
Herzlia 46661 Israel
Tel: +972-99577888
Fax: +972-99568860
eli@fly-supply.net
www.fly-supply.net
www.aircraft-partsupply.com

ITALY**FAE s.r.l.**

Via Tertulliano, 41
20137 Milano, Italy
Tel: +39-02-55187133
Fax: +39-02-55187399
fae@fae.it
www.fae.it

JAPAN**Tokyo Instruments, Inc.**

6-18-14 Nishikasai, Edogawa-ku,
Tokyo, 134-0088 Japan
Tel: +81 3 3686 4711
Fax: +81 3 3686 0831
f_kurabayashi@tokyoinst.co.jp
www.tokyoinst.co.jp

LATVIA**SIA "SOLARTEX"**

RAILWAY INSTRUMENTS ONLY
Duntes 15a, 5th floor, office B7
Riga, Latvia
Tel: +371 67 130 787
solartex@inbox.lv

MALAYSIA**OptoCom InstruVentures**

H-49-2, Jalan 5, Cosmoplex
Industrial Park, Bandar Baru
Salak Tinggi, Sepang, Malaysia
Tel: 603 8706 6806
Fax: 603 8706 6809
optocom@tm.net.my
www.optocom.com.my

MONTENEGRO**MTL ASCO Sp. z o.o.**

RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120
Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

NORWAY**BLConsult**

Ryssbalt 294,
95 291 Kalix, Sweden
Tel: +46 70 663 19 25
info@blconsult.se
www.blconsult.se

NORWAY**Salitec AS**

PB 468, N-1327
Lysaker, Norway
Tel: +47 23 891015
Fax: +47 92101005
mail@salitec.no
www.salitec.no

PERU
Verne Peru S.A.C.

Las Codornices 104,
Surquillo, Lima, Peru
Tel/fax: +51 992436734
info@verne.cl
www.verne.cl

POLAND
MTL ASCO Sp. z o.o.

RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120
Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

PORTUGAL
Campal Inovacoes

Ferroviarias Lda.
RAILWAY INSTRUMENTS ONLY
Lagoas Park, Edificio 7, 1º Piso Sul
2740-244 Porto Salvo,
Oeiras, Portugal
Tel: +351 21 584 4348
campal@campal.pt
www.campal.pt

**SERBIA, SLOVAKIA,
SLOVENIA**
MTL ASCO Sp. z o.o.

RAILWAY INSTRUMENTS ONLY
ul. Wielowiejska 53, 44-120
Pyskowice, Poland
Tel: + 48 32 332 70 03
Fax: + 48 32 332 70 14
rail@ascorail.eu
www.ascorail.eu

SOUTH KOREA
BS Holdings

B-201,Wonpogongwon 1ro,
59 Danwon-gu, Ansan-si,
Gyeonggi-do, 15455,
Republic of Korea
Tel: +82-31-411-5011
Fax: +82-31-411-5015
bsh5011@hanmail.net
www.lasersolution.co.kr

SOUTH KOREA
PROSEN. CO., LTD

M-1001, Songdo techno park IT
center, 32, Songdogwahak-ro,
Yeonsu-gu, Incheon, 21984,
Republic of Korea
Tel: +82-32-811-3457
Fax: +82-32-232-7458
trade@prosen.co.kr
www.prosen.co.kr

12
SPAIN
Iberfluid Instruments S.A.

C/ Botanica, 122, 08908
L'Hospitalet de Llobregat
Barcelona
Tel: +34 93 447 10 65
Fax: +34 93 334 05 24
myct@iberfluid.com
www.iberfluid.com

SWEDEN
BLConsult

Ryssbalt 294,
95 291 Kalix, Sweden
Tel: +46 70 663 19 25
info@blconsult.se
www.blconsult.se

SWEDEN
Kvalitest Industrial AB

**EXCEPT FOR RAILWAY
INSTRUMENTS**
Ekbacksaven 28,
16869 Bromma, Sweden
Tel: +46 0 76 525 5000
sales@kvalitest.com
www.kvalitest.com
www.kvalitest.se

SWITZERLAND
ID&T GmbH

Gewerbestrasse 12/a
8132 Egg (Zurich), Switzerland
Tel: + 41 44 994 92 32
Fax: + 41 44 994 92 34
info@idtlaser.com
www.idtlaser.com

THAILAND
Advantech Solution Co.,Ltd.

20/170 Motorway Rd.,
Kwang Pravet, Khet Pravet,
Bangkok, Thailand 10250
Tel: +662-1848705
Fax: +662-1848708
sales@advantechsolution.com
www.advantechsolution.com

TURKEY
MAK Elektronik Malzeme

**Analiz ve Kalite Kontrol
Cihazlari Dis Tic. Ltd. Sti.**
Cenap Sahabettin Sokak, No:39,
34718 Kosuyolu - Kadikoy /
Istanbul - TURKEY
Tel: +90 216 402 10 34
Fax: +90 216 402 10 35
ulastac@metalografi.net
www.makelektronik.com.tr

TURKEY
TEKMA Muhendislik A.S.

Cevizli Mh. M. Kemal Cd.,
Hukukcular Towers,
A-Blok, No: 66-A/39
Kartal - Istanbul
Tel: +90 216 970 1318
Tel: +90 850 840 2334
info@tekma.eu
www.tekma.eu

UKRAINE
KODA

Frunze st. 22, 61002,
Harkov, Ukraine
Tel/Fax: +38 057 714 26 54
mail@koda.com.ua
www.koda.com.ua

**UNITED KINGDOM,
IRELAND**
ALTHEN UK

Northamptonshire
United Kingdom
Tel: +44 (0) 7823 921427
t.stoyles@althen.co.uk
www.althensensors.com
www.althencontrols.com

USA, CANADA

Althen Sensors and Controls
2531 Bradley St.,
Oceanside, CA, 92056, USA
Tel: 858 633 3572
r.ream@althensensors.com

USA, CANADA, MEXICO

Acuity Products of Schmitt Industries, Inc.
2765 NW Nicolai Street
Portland, OR, 97210, USA
Tel: +1 503 227 7908
Fax: +1 503 223 1258
sales@acuitylaser.com
www.acuitylaser.com

USA, CANADA, MEXICO

International Electronic Machines Corporation
RAILWAY INSTRUMENTS ONLY
850 River Street, Troy,
New York, USA
Tel: +1 518 268-1636
Fax: +1 518 268-1639
marketing@iem.net
wwwиемnet